



Republic of the Philippines
Department of Health
CENTRAL OFFICE BIDS AND AWARDS COMMITTEE

BID BULLETIN NO. 3
07 December 2023

PROCUREMENT OF NEBULIZER AND BAG VALVE MASK
IB NO. 2024-109

This Bid Bulletin is being issued to respond to the queries during the Pre-Bidding Conference, to respond to the letter of the Prospective Bidder (PB), and to amend or modify the bidding documents posted on the PhilGEPS and DOH websites for the above-cited project. This Bid Bulletin shall form an integral part of the bidding document. Listed below are the corresponding modifications/changes:

1. Query from the PBs:

Query from the PBs		End-User Unit's (EUU's) Response
EUU's Specifications	Query	
<p>Detailed Technical Specifications:</p> <p>Item No.1 - Nebulizer</p> <p>4. Condensation rate: more than or equal to 0.2ml/min</p> <p>5. Atomized particles: less than or equal to 5um</p> <p>6. Residual liquid amount: less than 0.5ml</p> <p>7. Medication flow rate: approx. 0.35ml/min</p> <p>10. Protection level: IPX3</p> <p>11. Power supply: 2 AA batteries or DC 5V USB cable</p> <p>13. Package Inclusion: Extra two (2) AA battery</p>	<p>Vmed Medical Co. (VMC) requesting to change the following:</p> <p>Item No.1 - Nebulizer</p> <p>4. Condensation rate/Nebulization rate: >0.35ml/min</p> <p>5. Atomized particles/MMAD: less than 3um</p> <p>6. To delete</p> <p>7. To delete</p> <p>10. Protection level: at least IPX3</p> <p>11. To delete 2 AA batteries.</p> <p>13. To delete</p>	<p>Item No.1 - Nebulizer</p> <p>Not acceptable, within the range of our specification.</p> <p>Not acceptable, within the range of our specification.</p> <p>Not acceptable, no changes.</p> <p>Not acceptable, no changes.</p> <p>Acceptable, refer to changes.</p> <p>Refer to changes.</p> <p>Not acceptable, no changes.</p>

5. Section VII. Technical Specifications

From	To
<p>A. Detailed Technical Specifications:</p> <ol style="list-style-type: none"> 1. Mesh Nebulizer 2. Size: Manufacturer Standard 3. Material: ABS 4. Condensation Rate: more than or equal to 0.2ml/min. 5. Atomized Particles: less than or equal to 5µm 6. Residual Liquid Amount: less than 0.5ml 7. Medication flow rate: approx. 0.35ml/min 8. Cup capacity: at least 8ml 9. Noise: at least 50Db(A) 10. Protection level: IPX3 11. Power supply: 2 AA batteries or DC 5V USB cable 12. Portable 13. Package Inclusion: <ul style="list-style-type: none"> -1x Nebulizer -1x mask for adult -1x mask for child -1x Piece mouth -1x USB Cable -1x Instruction Manual -Extra two (2) AA Batteries 	<p>A. Detailed Technical Specifications:</p> <ol style="list-style-type: none"> 1. Mesh Nebulizer 2. Size: Manufacturer Standard 3. Material: ABS 4. Condensation Rate: more than or equal to 0.2ml/min. 5. Atomized Particles: less than or equal to 5µm 6. Residual Liquid Amount: less than 0.5ml 7. Medication flow rate: approx. 0.35ml/min 8. Cup capacity: at least 8ml 9. Noise: at least 50Db(A) 10. Protection level: <i>at least</i> IPX3 11. Power supply: 2 AA batteries <i>and</i> DC 5V USB cable 12. Portable 13. Package Inclusion: <ul style="list-style-type: none"> -1x Nebulizer -1x mask for adult -1x mask for child -1x Piece mouth -1x USB Cable -1x Instruction Manual -Extra two (2) AA Batteries

The revised Technical Specifications is enclosed for the Prospective Bidder’s reference and use.

All other provisions indicated in the bidding documents which are not affected shall remain in force and in effect.

For guidance and information of all concerned.

KENNETH G. RONQUILLO, MD, MPH, CESO III
 Undersecretary of Health
 Chairperson, COBAC – E

Technical Specifications

Republic of the Philippines
Department of Health
TECHNICAL SPECIFICATIONS

Item No. 1	Nebulizer	Quantity / Unit	2,000 units
Name of Manufacturer:		Country of Origin:	
Brand:			
Total ABC for Item No. 1: PhP2,000,000.00			
PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
A. <u>Detailed Technical Specifications:</u> <ol style="list-style-type: none"> 1. Mesh Nebulizer 2. Size: Manufacturer Standard 3. Material: ABS 4. Condensation Rate: more than or equal to 0.2ml/min. 5. Atomized Particles: less than or equal to 5µm 6. Residual Liquid Amount: less than 0.5ml 7. Medication flow rate: approx. 0.35ml/min 8. Cup capacity: at least 8ml 9. Noise: at least 50Db(A) 10. Protection level: <i>at least</i> IPX3 11. Power supply: 2 AA batteries <i>and</i> DC 5V USB cable 12. Portable 13. Package Inclusion: <ul style="list-style-type: none"> -1x Nebulizer -1x mask for adult -1x mask for child -1x Piece mouth -1x USB Cable -1x Instruction Manual -Extra two (2) AA Batteries 			
B. <u>Additional Requirements to be attached to Technical Specifications form arranged, numbered and tabbed as enumerated below:</u> <ol style="list-style-type: none"> 1. Valid and current Certificate of Medical Device Registration (CMDR) or Certificate of Medical Device Notification (CMDN) issued by Philippine Food and Drug Administration (PFDA); 2. Valid and current License to Operate (LTO) for Medical Device Importer/Wholesaler issued by PFDA; 3. The bidder shall submit any of the following whichever is applicable: <ol style="list-style-type: none"> a. If the bidder is a manufacturer, certificate that the bidder manufactures the products/item; or 			

Republic of the Philippines
Department of Health
TECHNICAL SPECIFICATIONS

Item No. 1	Nebulizer	Quantity / Unit	2,000 units
Name of Manufacturer:		Country of Origin:	
Brand:			
Total ABC for Item No. 1: PhP2,000,000.00			
PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
<p>b. If the bidder is an Exclusive/Authorized Distributor or Dealer of the products/items, a Certificate or Contract from the manufacturer must be provided as proof that the bidder is an Exclusive/Authorized Distributor or Dealer of the products/items; or</p> <p>c. If the bidder is an agent of the exclusive distributor or dealer, the following must be provided:</p> <p style="margin-left: 40px;">i. Certificate or Distributor/Dealership Agreement by the Manufacturer with the distributor or dealer; and</p> <p style="margin-left: 40px;">ii. Certificate or Contract/ Dealership Agreement between the distributor/dealer and the bidder.</p> <p>4. Product Insert/Product Information or downloaded from the internet and other manufacturer's un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate for cross-referencing statement of compliance to the technical specification in accordance to what is indicated in 2nd page of Section VII. Technical Specifications of the Bidding Documents;</p> <p>5. Valid and Current Declaration of Conformity with appropriate ISO/IEC/PNS Standard issued to the manufacturer by an ISO Certifying body or any appropriate agency or body;</p> <p>6. Warranty Certificate for One (1) year on parts and services;</p> <p>7. Certificate that the Product brand has been in local market for at least 3 years;</p> <p>8. Sworn Statement using the prescribed form.</p>			
C. <u>Additional Requirements to be submitted by the Lowest/Single Calculated Bidder (S/LCB):</u>			
1. Submit one (1) sample, for evaluation.			
D. <u>Upon delivery the following shall be complied with:</u>			
1. Packaging instructions: Standard packaging of the manufacturer			
2. Labeling instructions:			
<p style="margin-left: 40px;">i. On each box and unit, the following shall be imprinted or stickered with non-removable or permanent sticker or label that is binding and with residue and tearing if removed:</p>			
<p>Philippine Government Property – Department of Health NOT FOR SALE</p>			
E. <u>Product Recall & Disposal:</u>			
1. The Supplier must ensure the quality of products and if there will be problems in the quality, the Supplier will recall and replace the products distributed in the regions/hospitals/treatment			

Republic of the Philippines
Department of Health
TECHNICAL SPECIFICATIONS

Item No. 1	Nebulizer	Quantity / Unit	2,000 units
Name of Manufacturer:		Country of Origin:	
Brand:			
Total ABC for Item No. 1: PhP2,000,000.00			
PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
hubs/RHU/HC/BHSS based on Guidelines on Product Recall, FDA Circular No. 2016-012; 2. In case of recalls, or damage due to replacement, the costs associated with the proper handling or pull out from health facilities where the supplies or goods have already been distributed shall be borne by the Supplier (DOH Administrative Order (AO) No. 2019-0041).			

Signature over Printed Name

[date of signing]

In the capacity of:

Duly authorized to sign bid for and on behalf of:

[title or other appropriate designation]

[Name of Company]

[Complete office address]

[Contact No.]

[Fax No.]